














MASTER COPY

	NOVO EXCIPIENTS PVT. LTD., NAVI MUMBAI WAREHOUSE DEPARTMENT		Issued By/ On:	 26/09/23
			No. of Copies:	3/3
Format Title :		Daily Verification Of Weighing Balance (600 gm.)		
Format No.:		F/SOP/WR/019/01-01	Page No.:	1 of 2
Ref. SOP No.:		SOP/WR/019		

Department	Warehouse	Balance Location	Dispensing room for powder
ID No.	WR-002	Capacity (a)	600.00 g
		Least Count (c)	0.01 g

Standard Weights		Tolerance	Accuracy Limits
Minimum Weight (In house)	0.500 g	± 0.1 %	0.4995 To 0.5005 g
Middle Weight (50% of a)	300.00 g	± 0.1 %	299.700 To 300.300 g
Maximum Weight (In house)	500.00 g	± 0.1 %	499.500 To 500.500 g

Date	Standard Weight			Done By	Checked By / On	Remark
	0.500 g	300.000 g	500.000 g			
01/10/23	Weekly off					
02/10/23	0.500g	300.010g	500.010g	Mahesh	 02/10/23	ok
03/10/23	0.500g	300.000g	500.010g	Mahesh	 03/10/23	ok
04/10/23	0.500g	300.010g	500.020g	Mahesh	 04/10/23	ok
05/10/23	0.500g	300.000g	500.000g	Mahesh	 05/10/23	ok
06/10/23	Weekly off					
07/10/23	0.500g	300.000g	500.010g	Mahesh	 07/10/23	ok
08/10/23	0.500g	300.000g	500.000g	Mahesh	 08/10/23	ok
09/10/23	0.500g	300.000g	500.010g	Mahesh	 09/10/23	ok
10/10/23	0.500g	299.990g	499.990g	Mahesh	 10/10/23	ok
11/10/23	0.500g	299.990g	499.990g	Mahesh	 11/10/23	ok
12/10/23	0.500g	299.990g	500.010g	Mahesh	 12/10/23	ok

	PREPARED BY	REVIEWED BY	APPROVED BY
Name	Dattatray Khose	Pooja Yelwande	Karan Patil
Sign/Date	 09/10/23	 09/10/23	 10/10/23
Designation	Sr. officer	Sr. officer	Manager
Department	Warehouse	Warehouse	Quality Assurance